

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

QUALITY CONTROL CERTIFICATE

KIT NAME CHORUS PROMONITOR ANTI-INFLIXIMAB

Code	86701		
Lot No.	853	2023-07	2022-07

REAGENTS

Label	Description	Lot.	Exp	No
D4320	PROMONITOR ANTI-INFLIXIMAB	853	2023-07	6X6
B4320	CALIBRATOR ANTI-INFLIXIMAB	853	2024-05	1
C4320	CONTROL ANTI-INFLIXIMAB	853	2024-05	1

CONTROLS

Found results	Expected results
---------------	------------------

Calibrator (mOD)	444	≥ 100
Control (AU/ml)	22,9	21,6-34,5

CALIBRATOR

2	7	5	8	5	3	6	5	2	6	4	7	3	6	7	7	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader YES ☒ Operator Signature *SB*



CONTROL SERUM

2	7	5	8	5	3	6	2	3	3	4	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature *SB*



This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... *Alvisei*

DATE 05.08.2022

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.