

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI) - Italia - Via delle Rose 10
Tel +39-0577-587111 Fax +39-0577-318690



Pag 1/1
Sede Legale e Amministrazione
20144 Milano (Italy) Via A. Solari 19 sc. 6
Tel. +39-02-4859121 r.a. Fax +39-02-48008530

QUALITY CONTROL CERTIFICATE

KIT NAME: BETA 2-GLYCOPROTEIN-M - CHORUS

Code 86052	
Lot No. 836	2023-05 2022-05

REAGENTS

Label	Description	Lot.	Exp	No
D3371	BETA 2-GLYCOPROTEIN-M	836	2023-05	6X6
B3371	CALIBRATORE BETA 2-GLYCOPROTEIN-M	836	2023-05	1
C3371	CONTROL SERUM BETA 2-GLYCOPROTEIN-M	836	2023-05	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	320	≥ 100
Control Serum (AU/mL)	79,9	46,5-128,6

CALIBRATOR

2	0	5	8	3	6	6	3	3	0	4	9	2	5	7	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader

YES



Operator Signature

[Signature]



CONTROL SERUM

2	0	5	8	3	6	6	2	8	5	2	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature

[Signature]



The test has been carried out in accordance with the package insert

TESTED BY: Simone Baiocchi

[Signature]

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

[Signature]

DATE

24.08.2022

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.