

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SIENA) – Italy – Strada dei Laghi, 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale e Amministrazione
Via A. Solari, 19 20144 Milano - Italy

QUALITY CONTROL CERTIFICATE

KIT NAME ENZYWELL RUBELLA IgG

Code 91030

LOT.NO 280-A EXPIRY 2023-10

REAGENTS

CODE	LABEL	DESCRIPTION	LOT No.	EXP	No
PF93030	006510	MICROPLATE	201	2024-06	1
PF93507	006500	CONJUGATE	237	2023-10	1
PF93910	1036	IgG NEGATIVE CONTROL	173-2	2023-11	1
PF93831	006520	CALIBRATOR 1	197	2024-01	1
PF91830	006540	CALIBRATOR 2	197	2024-01	1
PF93833	006550	CALIBRATOR 3	197	2024-01	1
PF93834	006560	CALIBRATOR 4	197	2024-01	1
PF93835	006530	CALIBRATOR 5	197	2024-01	1
PF93619	1034	SUBSTRATE-HS	364-1	2023-11	1
PF93611	009660	DILUENT 2	224	2023-10	1
PF93603	1035	WASHING BUFFER 10x	450-1	2023-11	1
PF93602	1032/RISK0	STOP SOLUTION	214-4	2024-02	1

CONTROLS

IU/mL	FOUND RESULTS	EXPECTED RESULTS
IgG NEG CONTROL 0	0.010	≤ 0,6 CUT-OFF
CALIBRATOR 1 5	0.145	< CUT-OFF
CALIBRATOR 2 10 (cut-off)	0.374	≥ 0.200
CALIBRATOR 3 50	1.000	
CALIBRATOR 4 100	1.787	
CALIBRATOR 5 200	2.604	≥ 1.500
CV %	14%	≤ 15%

The test has been carried out in accordance with the package insert

TESTED BY: Simone Baiocchi *Simone Baiocchi*

APPROVED BY: (RESPONSIBLE HEAD, Quality Control Reagents) *[Signature]*

DATE: 25/08/2022

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested by an FDA approved method and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.



CERTIFICATE OF ANALYSIS

KIT:	RUBELLA IgG	Code 91030
LOT:	280	Code 91176
DATE OF TEST:	10/08/2022	

METHOD: ACCORDING TO THE PACKAGE INSERT

		<i>FOUND</i>	<i>EXPECTED</i>
1. SOLID PHASE			
A) UNIFORMITY BETWEEN PLATES			
1st PLATE	OD450 nm	2.572	
	CV%	5,7%	$\leq 20\%$
2nd PLATE	OD 450nm	2.715	
	CV%	3,7%	$\leq 20\%$
B) UNIFORMITY WITHIN PLATE			
	OD 450nm	2.514	
	CV%	4,9%	$\leq 15\%$
2. CONTROL OF COMPLETE KIT			
A) Substrate Blank	OD 450 nm	0.059	≤ 0.100
Calibrator 0 IU/ml	OD 450 nm	0.010	$\leq 0,6 \times \text{Cut-off}$
Calibrator 5 IU/ml	OD 450 nm	0.145	$< \text{Cut-off}$
Calibrator 10 IU/ml (Cut-off)	OD 450 nm	0.374	> 0.200
Calibrator 50 IU/ml	OD 450 nm	1.000	
Calibrator 100 IU/ml	OD 450 nm	1.787	
Calibrator 200 IU/ml	OD 450 nm	2.604	≥ 1.500
B) PANEL OF HUMAN SERA AND INTERNAL STANDARDS			
Results of Negative sera		Accepted	According to spec.
Results of Positive sera		Accepted	According to spec.
Results of Internal standards		Accepted	According to spec.
C) HOMOGENEITY OF CUT-OFF			
	No.	6	
	OD 450 nm	0.433	
	CV%	14%	$\leq 15\%$
D) PRECISION OF FILLING PROCESS			

☒ SATISFACTORY

☐ UNSATISFACTORY

FINAL RESULT: **ACCEPTED**

Date: **10/08/2022**

Signed: 
Released: Q.C.Manager

Aggiornamento 05.2012