

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME CHORUS PROMONITOR USTEKINUMAB

Code 86712

Lot No. 852



2023-07



2022-07

REAGENTS

Label	Description	Lot.	Exp.	NO
D4500	PROMONITOR USTEKINUMAB	852	2023-07	6x6
B4500	CALIBRATORE USTEKINUMAB	852	2024-05	1
C4500	CONTROL SERUM USTEKINUMAB	852	2024-05	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	209	≥ 100
Control Serum (ug/mL)	1,2	1,0-1,5

CALIBRATOR

2	8	0	8	5	2	6	2	2	3	4	3	5	5	6	5	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature



CONTROL SERUM


2	8	0	8	5	2	6	1	5	2	2	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.....

DATE 13/09/2022

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.