

**DIESSE DIAGNOSTICA SENESE SpA**  
53035 Monteriggioni (SI) – Italia – Via delle Rose 10  
Tel +39-0577-587111 Fax +39-0577-318690



Pag 1/1  
Sede Legale e Amministrazione  
20144 Milano (Italy) Via A. Solari 19 sc. 6  
Tel. +39-02-4859121 r.a. Fax +39-02-48008530

# QUALITY CONTROL CERTIFICATE

KIT NAME: DEAMIDATED GLIADIN-A - CHORUS

|             |         |         |
|-------------|---------|---------|
| Code 86098  |         |         |
| Lot No. 853 | 2023-07 | 2022-07 |

## REAGENTS

| Label | Description                        | Lot. | Exp     | No  |
|-------|------------------------------------|------|---------|-----|
| D8650 | Deamidated Gliadin-A               | 853  | 2023-07 | 6X6 |
| B8650 | CALIBRATORE Deamidated Gliadin-A   | 853  | 2023-07 | 1   |
| C8650 | CONTROL SERUM Deamidated Gliadin-A | 853  | 2023-07 | 1   |

## CONTROLS


| Found results | Expected results |
|---------------|------------------|
|---------------|------------------|

|                       |      |            |
|-----------------------|------|------------|
| Calibrator (mOD)      | 500  | ≥ 100      |
| Control Serum (AU/mL) | 72,3 | 38,3-150,0 |

## CALIBRATOR

|   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|
| 2 | 5 | 4 | 8 | 5 | 3 | 6 | 2 | 2 | 2 | 3 | 0 | 4 | 0 | 4 | 2 | 0 | 0 |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|

Label checked with barcode reader YES ☒ Operator Signature



## CONTROL SERUM

|   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|
| 2 | 5 | 4 | 8 | 5 | 3 | 6 | 1 | 8 | 3 | 3 | 0 | 0 | 0 | 0 | 0 | 0 | 1 |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|

Label checked with barcode reader YES ☒ Operator Signature



The test has been carried out in accordance with the package insert

TESTED BY: Chiara Bruschelli   
APPROVED RESPONSIBLE HEAD, Quality Control Reagents   
DATE 12.09.2022

## WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested by an FDA approved method and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.