

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI) - Italia - Via delle Rose 10
Tel +39-0577-587111 Fax +39-0577-318690



Pag 1/1
Sede Legale e Amministrazione
20144 Milano (Italy) Via A. Solari 19 sc. 6
Tel. +39-02-4859121 r.a. Fax +39-02-48008530

QUALITY CONTROL CERTIFICATE

KIT NAME:CHIKUNGUNYA IgG -CHORUS

Code 81133

Lot No. 852



2023-07



2022-07

REAGENTS

Label	Description	Lot.	Exp	No
D0200	CHIKUNGUNYA IgG	852	2023-07	6X6
B0200	CALIBRATORE CHIKUNGUNYA IgG	852	2023-07	1
C0200	CONTROL CHIKUNGUNYA IgG	852	2023-07	1

CONTROLS

	Found results	Expected result
Calibrator (mOD)	608	≥ 100
Control (INDEX)	1,9	1,3-2,3

CALIBRATOR


9	0	9	8	5	2	6	0	0	0	0	0	5	6	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature



CONTROL +

9	0	9	8	5	2	6	1	3	2	3	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature



The test has been carried out in accordance with the package insert

TESTED BY: Chiara Bruschelli

APPROVED

RESPONSIBLE HEAD, Quality Control Reagents.....

DATE

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.