

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI) – Italia – Via delle Rose 10
Tel +39-0577-587111 Fax +39-0577-318690



Pag 1/1

Sede Legale e Amministrazione
20144 Milano (Italy) Via A. Solari 19 sc. 6
Tel. +39-02-4859121 r.a. Fax +39-02-48008530

QUALITY CONTROL CERTIFICATE

KIT NAME: tTg-G - CHORUS

Code	86060		
Lot No.	871	2023-09	2022-09

REAGENTS

Label	Description	Lot.	Exp	No
D3451	tTg-G	871	2023-09	6X6
B3451	CALIBRATORE tTg-G	871	2023-09	1
C3451	CONTROL SERUM tTg-G	871	2023-09	1

CONTROLS


	Found results	Expected results
Calibrator (mOD)	403	≥ 100
Control Serum (AU/mL)	36,2	25,0-52,0

CALIBRATOR

3	4	0	8	7	1	6	1	1	9	3	4	4	4	4	2	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES


Operator Signature *te*

CONTROL SERUM

3	4	0	8	7	1	6	1	6	3	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES

Operator Signature *te*

The test has been carried out in accordance with the package insert

TESTED BY: Antonella Leo *Antonella Leo*APPROVED RESPONSIBLE HEAD, Quality Control Reagents *Bruno*DATE *27/09/2024*

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested by an FDA approved method and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.