

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI) – Italia – Via delle Rose 10
Tel +39-0577-587111 Fax +39-0577-318690



Pag 1/1
Sede Legale e Amministrazione
20144 Milano (Italy) Via A. Solari 19 sc. 6
Tel. +39-02-4859121 r.a. Fax +39-02-48008530

QUALITY CONTROL CERTIFICATE

KIT NAME: BETA 2-GLYCOPROTEIN-M - CHORUS

Code 86052		
Lot No. 864	2023-08	2022-09

REAGENTS

Label	Description	Lot.	Exp	No
D3371	BETA 2-GLYCOPROTEIN-M	864	2023-08	6X6
B3371	CALIBRATORE BETA 2-GLYCOPROTEIN-M	864	2023-09	1
C3371	CONTROL SERUM BETA 2-GLYCOPROTEIN-M	864	2023-09	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	290	≥ 100
Control Serum (AU/mL)	75,6	41,7-108,0


CALIBRATOR

2	0	5	8	6	4	6	4	3	1	4	7	2	7	7	2	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES ☒

Operator Signature *SS*




CONTROL SERUM

2	0	5	8	6	4	6	2	6	4	8	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES ☒

Operator Signature *SS*



The test has been carried out in accordance with the package insert

TESTED BY: Antonella Leo *Antonella Leo*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *dfw*

DATE 28-09-2022

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.