

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI) - Italia - Via delle Rose 10
Tel +39-0577-587111 Fax +39-0577-318690



Pag 1/1
Sede Legale e Amministrazione
20144 Milano (Italy) Via A. Solari 19 sc. 6
Tel. +39-02-4859121 r.a. Fax +39-02-48008530

QUALITY CONTROL CERTIFICATE

KIT NAME: TSH

Code	86516		
Lot No.	869	2023-09	2022-09

REAGENTS

Label	Description	Lot.	Exp	No
D3290	TSH	869	2023-09	6X6
B3290	CALIBRATOR TSH	869	2023-09	1
C3290	CONTROL SERUM TSH	869	2023-09	1

CONTROLS

	Found results	Expected results
CALIBRATOR (mOD)	310	≥100
CONTROL SERUM (uIU/ml)	4,3	2,1-6,8

CALIBRATOR


9	1	9	8	6	9	6	7	1	4	1	8	4	7	3	9	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature



CONTROL SERUM


9	1	9	8	6	9	6	1	5	2	8	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature



The test has been carried out in accordance with the package insert

TESTED BY: Sara Benocci

APPROVED

RESPONSIBLE HEAD, Quality Control Reagents.....

DATE

04.10.2022

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.