

QUALITY CONTROL CERTIFICATE

KIT NAME BETA 2-GLYCOPROTEIN-M - CHORUS

Code 86052

Lot No. 865



2023-08



2022-09

REAGENTS

Label	Description	Lot.	Exp.	NO
D3371	BETA 2-GLYCOPROTEIN-M	865	2023-08	6X6
B3371	CALIBRATORE BETA 2-GLYCOPROTEIN-M	865	2023-09	1
C3371	CONTROL SERUM BETA 2-GLYCOPROTEIN-M	865	2023-09	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	389	> 100
Control Serum (AU/mL)	49,7	41,7-108,0

CALIBRATOR


2	0	5	8	6	5	6	4	3	1	4	7	2	7	7	2	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature



CONTROL SERUM

2	0	5	8	6	5	6	2	6	4	8	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.....

DATE 06-10-2022

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.