

DIESE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME **Scl-70 - CHORUS**

Code **86026**

Lot No. **880**



2023-10



2022-10

REAGENTS

Label	Description	Lot.	Exp.	NO
D3221	Scl-70	880	2023-10	6X6
B3221	CALIBRATORE Scl-70	880	2023-10	1
C3221	CONTROL SERUM Scl-70	880	2023-10	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	354	≥ 100
Control Serum (AU/mL)	57,7	36,9-86,7

CALIBRATOR


1	9	5	8	8	0	6	2	2	0	3	3	3	0	4	8	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature *SB*



CONTROL SERUM

1	9	5	8	8	0	6	2	3	4	3	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature *SB*



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... *SB*

DATE *20-10-2022*

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.