

DIESSE DIAGNOSTICA SENESE SpA  
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39  
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale  
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

## QUALITY CONTROL CERTIFICATE

KIT NAME tTg-A - CHORUS

Code 86058

Lot No. 872



2023-09



2022-09

### REAGENTS

Label	Description	Lot.	Exp.	NO
D3431	tTg-A	872	2023-09	6X6
B3431	CALIBRATORE tTg-A	872	2023-09	1
C3431	CONTROL SERUM tTg-A	872	2023-09	1

### CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	510	$\geq 100$
Control Serum (AU/mL)	23,0	19,1-36,0

### CALIBRATOR


3	3	9	8	7	2	6	3	2	1	3	6	5	5	4	4	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature .....



### CONTROL SERUM

3	3	9	8	7	2	6	1	3	2	4	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature .....



*This batch has been manufactured and checked according to quality requirements.  
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.....

DATE 18.10.2022

### WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.