

DIESSE DIAGNOSTICA SENESE SpA  
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39  
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale  
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

## QUALITY CONTROL CERTIFICATE

KIT NAME SS-B - CHORUS

Code 86024

Lot No. 878



2023-10



2022-10

### REAGENTS

Label	Description	Lot.	Exp.	NO
D3201	SS-B	878	2023-10	6X6
B3201	CALIBRATORE SS-B	878	2023-10	1
C3201	CONTROL SERUM SS-B	878	2023-10	1

### CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	380	$\geq 100$
Control Serum (AU/mL)	46,8	34,7-61,6

### CALIBRATOR

1	9	4	8	7	8	6	1	1	7	3	6	3	5	6	1	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader

YES



Operator Signature

*[Signature]*



### CONTROL SERUM

1	9	4	8	7	8	6	2	0	3	7	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature

*[Signature]*



*This batch has been manufactured and checked according to quality requirements.  
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

*[Signature]*

DATE 26-10-2022

### WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.