

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI) – Italia – Via delle Rose 10
Tel +39-0577-587111 Fax +39-0577-318690



Pag 1/1
Sede Legale e Amministrazione
20144 Milano (Italy) Via A. Solari 19 sc. 6
Tel. +39-02-4859121 r.a. Fax +39-02-48008530

QUALITY CONTROL CERTIFICATE

KIT NAME: INTRINSIC FACTOR - CHORUS

Code	86090		
Lot No.	878	2023-10	2022-10

REAGENTS

Label	Description	Lot.	Exp	No
D7431	INTRINSIC FACTOR	878	2023-10	6X6
B7431	CALIBRATORE INTRINSIC FACTOR	878	2023-10	1
C7431	CONTROL SERUM INTRINSIC FACTOR	878	2023-10	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	390	≥ 100
Control Serum (AU/mL)	90,1	60,7-175,0

CALIBRATOR


5	1	2	8	7	8	6	2	1	7	3	1	3	5	4	7	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature ... *SB*



CONTROL SERUM


5	1	2	8	7	8	6	2	0	3	7	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature ... *SB*



The test has been carried out in accordance with the package insert

TESTED BY: Simone Baiocchi

Simone Baiocchi

APPROVED

RESPONSIBLE HEAD, Quality Control Reagents

[Signature]

DATE

28.10.2021

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested by an FDA approved method and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.