DIESSE DIAGNOSTICA SENESE SpA

53035 Monteriggioni (SI), Italy, Strada dei Laghi 39 Tel. +39 0577319554 FAX +39 0577319020



Sede Legale 20144 Milano (Italy) Via A. Solari 19 sc. 6

QUALITY CONTROL CERTIFICATE KIT NAME TRAB - CHORUS Code 86112 Lot No. 878 ∑ 2023-07 2022-10

REAGENTS

Label	Description	Lot.	Exp.	NO
D4041	TRAB	878	2023-07	6X6
B4040	CALIBRATORE TRAB	878	2024-01	1.
C4040	CONTROL SERUM TRAB	878	2024-01	1

CONTROLS

Found results	Expected results

Calibrator (mOD)	371	≥ 100
Control Serum (U/L)	2,1	0,40-5,2

CALIBRATOR

9 4 7 8 7 8 5 4 2 1 2 9 4 2 3 1 0 0

Label checked with barcode reader

YES X

Operator Signature



CONTROL SERUM

9 4 7 8 7 8 5 1 7 3 1 0 0 0 0 0 1

Label checked with barcode reader

YES 🔀

Operator Signature ..



This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.......

Pierel.

DATE 07-11-2012

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.