

DIESSE DIAGNOSTICA SENESE SpA  
53035 Monteriggioni (SI) - Italia - Via delle Rose 10  
Tel +39-0577-587111 Fax +39-0577-318690



Pag 1/1  
Sede Legale e Amministrazione  
20144 Milano (Italy) Via A. Solari 19 sc. 6  
Tel. +39-02-4859121 r.a. Fax +39-02-48008530

# QUALITY CONTROL CERTIFICATE

KIT NAME: 88000-G6 TIMOTHY GRASS

|         |          |         |         |
|---------|----------|---------|---------|
| Code    | 88000-G6 |         |         |
| Lot No. | 870      | 2023-09 | 2022-11 |

## REAGENTS

| Label | Description                          | Lot | Exp     | No  |
|-------|--------------------------------------|-----|---------|-----|
| D9911 | 88000-G6 TIMOTHY GRASS               | 870 | 2023-09 | 2X6 |
| B9910 | CALIBRATORE 88000-G6 TIMOTHY GRASS   | 870 | 2023-10 | 1   |
| C9910 | CONTROL SERUM 88000-G6 TIMOTHY GRASS | 870 | 2023-10 | 1   |


## CONTROLS

|                      | Found results | Expected result |
|----------------------|---------------|-----------------|
| Calibrator (mOD)     | 270           | $\geq 100$      |
| Control Serum (KU/L) | 14,0          | 6,0-22,8        |

## CALIBRATOR

|   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|
| 6 | 1 | 2 | 8 | 7 | 0 | 6 | 5 | 2 | 7 | 5 | 7 | 2 | 2 | 7 | 7 | 0 | 0 |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|


Label checked with barcode reader YES ☒ Operator Signature *SB*



## CONTROL SERUM

|   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|
| 6 | 1 | 2 | 8 | 7 | 0 | 6 | 3 | 2 | 6 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 1 |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|

Label checked with barcode reader YES ☒ Operator Signature *SB*



The test has been carried out in accordance with the package insert

TESTED BY: Simone Baiocchi *Simone Baiocchi*  
APPROVED RESPONSIBLE HEAD, Quality Control Reagents. *[Signature]*  
DATE *08.11.2022*

## WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested by an FDA approved method and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.