

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME DEAMIDATED GLIADIN-A - CHORUS

Code 86098

Lot No. 882



2023-10



2022-11

REAGENTS

Label	Description	Lot.	Exp.	NO
D8650	Deamidated Gliadin-A	882	2023-10	6X6
B8650	CALIBRATORE Deamidated Gliadin-A	882	2023-10	1
C8650	CONTROL SERUM Deamidated Gliadin-A	882	2023-10	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	420	≥ 100
Control Serum (AU/mL)	56,1	30,9-75,0

CALIBRATOR

2	5	4	8	8	2	6	2	2	1	3	3	4	8	4	6	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒

Operator Signature 




CONTROL SERUM

2	5	4	8	8	2	6	1	5	2	7	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒

Operator Signature 



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents 

DATE 15.11.2022

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.