

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI) - Italia - Via delle Rose 10
Tel +39-0577-587111 Fax +39-0577-318690



Sede Legale e Amministrazione
20144 Milano (Italy) Via A. Solari 19 sc. 6
Tel. +39-02-4859121 r.a. Fax +39-02-48008530

QUALITY CONTROL CERTIFICATE

KIT NAME: FT4 - CHORUS

Code	86520		
Lot No.	887	2023-11	2022-11

REAGENTS

Label	Description	Lot.	Exp	No
D3320	FT4	887	2023-11	6X6
B3320	CALIBRATOR FT4	887	2023-11	1
C3320	CONTROL SERUM FT4	887	2023-11	1

CONTROLS

	Found results	Expected results
CALIBRATOR (mOD)	350	≥100
CONTROL SERUM (pmol/L)	26,0	9,8-31,8

CALIBRATOR


9	6	8	8	8	7	6	2	2	9	3	9	3	1	4	9	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature



CONTROL SERUM


9	6	8	8	8	7	6	2	3	4	2	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature



The test has been carried out in accordance with the package insert

TESTED BY: Antonella Leo

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.....

DATE 18.11.2022

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.