

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME Epstein-Barr VCA IgM II - CHORUS

Code 81054

Lot No. 887



2024-02



2022-11

REAGENTS

Label	Description	Lot.	Exp.	NO
D9490	Epstein-Barr VCA IgM II	887	2024-02	6x6
B9490	CALIBRATOR Epstein-Barr VCA IgM II	887	2024-02	1
C9490	CONTROL SERUM Epstein-Barr VCA IgM II	887	2024-02	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	458	≥ 100
Control Serum (INDEX)	2,1	1,4-2,6
Negative Control (INDEX)	0,1	$< 0,7$

CALIBRATOR

0	3	0	8	8	7	7	0	0	0	0	0	5	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature



CONTROL SERUM

0	3	0	8	8	7	7	1	4	2	6	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.....

DATE 26.11.2022

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.