

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI) – Italia – Via delle Rose 10
Tel +39-0577-587111 Fax +39-0577-318690



Pag 1/1
Sede Legale e Amministrazione
20144 Milano (Italy) Via A. Solari 19 sc. 6
Tel. +39-02-4859121 r.a. Fax +39-02-48008530

QUALITY CONTROL CERTIFICATE

KIT NAME: 88000-C286 - IBUPROFEN

Code	88000 - C286		
Lot No.	852	2023-07	2022-08

REAGENTS

Label	Description	Lot.	Exp	No
D5841	88000- C286 IBUPROFEN	852	2023-07	2X6
B5840	CALIBRATORE 88000- C286 IBUPROFEN	852	2023-10	1
C5840	CONTROL SERUM 88000- C286 IBUPROFEN	852	2023-10	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	270	≥ 100
Control Serum (KU/L)	14,6	6,0-22,8

CALIBRATOR


7	2	1	8	5	2	6	5	2	7	5	7	2	2	7	7	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature*SB*.....



CONTROL SERUM


7	2	1	8	5	2	6	3	2	6	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature*SB*.....



The test has been carried out in accordance with the package insert

TESTED BY: Simone Balocchi *Simone Balocchi*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.....*[Signature]*.....

DATE *26.08.2022*

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested by an FDA approved method and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.