

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME MYCOPLASMA PNEUMONIAE IgM - CHORUS

Code 81035

Lot No. 889



2024-02



2022-11

REAGENTS

Label	Description	Lot.	Exp.	NO
D2261	MYCOPLASMA PNEUMONIAE IgM	889	2024-02	6x6
B2261	CALIBRATORE MYCOPLASMA PNEUMONIAE IgM	889	2024-02	1
C2261	CONTROL SERUM MYCOPLASMA PNEUMONIAE IgM	889	2024-02	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	470	≥ 100
Control Serum (INDEX)	2,1	1,4-2,5
Negative Control (INDEX)	0,0	$\leq 0,7$


CALIBRATOR

0	6	7	8	8	9	7	0	0	0	0	0	5	2	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES ☒

Operator Signature



CONTROL SERUM

0	6	7	8	8	9	7	1	4	2	5	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES ☒

Operator Signature



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.....

DATE 06-12-2022

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.