

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME Vitamin D Total- CHORUS

Code 86900

Lot No. 871



2023-09



2022-09

REAGENTS

Label	Description	Lot.	Exp.	NO
D9702	25OH Vitamin D Total	871	2023-09	6x6
B9703	CALIBRATOR 25OH Vitamin D Total	871	2024-09	1
C9703	CONTROL SERUM 25OH Vitamin D Total	871	2024-09	4

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	586	≥ 100
CONTROL SERUM (ng/ml)	35,1	30,3-37,9

CALIBRATOR

8	3	5	8	7	1	6	2	3	6	3	8	5	5	5	7	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature



CONTROL SERUM

8	3	5	8	7	1	6	1	3	2	4	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.....

DATE 07-12-2022

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent reagents should be handled with maximum attention.