

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME Jo-1 - CHORUS

Code	86030
Lot No.	887
	2024-02
	2022-11

REAGENTS

Label	Description	Lot.	Exp.	NO
D3261	Jo-1	887	2024-02	6X6
B3261	CALIBRATORE Jo-1	887	2024-02	1
C3261	CONTROL SERUM Jo-1	887	2024-02	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	362	≥ 100
Control Serum (AU/mL)	58,9	33,0-100,0

CALIBRATOR

1	9	7	8	8	7	7	1	1	9	2	9	3	6	3	6	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature



CONTROL SERUM

1	9	7	8	8	7	7	2	0	3	6	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.....

DATE 12-12-2022

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.