

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME PR3 - CHORUS

Code 86066

Lot No. 889



2023-11



2022-11

REAGENTS

Label	Description	Lot.	Exp.	NO
D3512	PR3	889	2023-11	6X6
B3512	CALIBRATORE PR3	889	2023-11	1
C3512	CONTROL SERUM PR3	889	2023-11	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	450	≥ 100
Control Serum (AU/mL)	70,7	50,0-97,5

CALIBRATOR

2	5	2	8	8	9	6	5	1	3	2	5	3	3	4	1	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES ☒

Operator Signature 




CONTROL SERUM

2	5	2	8	8	9	6	2	1	4	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES ☒

Operator Signature 



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents 

DATE 17.12.22

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.