

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME dsDNA-M - CHORUS

Code 86034

Lot No. 879



2024-01



2022-10

REAGENTS

Label	Description	Lot.	Exp.	NO
D7401	dsDNA-M	879	2024-01	6X6
B7401	CALIBRATORE dsDNA-M	879	2024-01	1
C7401	CONTROL. SERUM dsDNA-M	879	2024-01	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	624	≥ 100
Control Serum (AU/mL)	50,0	30,0-69,2

CALIBRATOR


2	5	7	8	7	9	7	6	1	5	2	8	6	6	4	1	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature



CONTROL SERUM


2	5	7	8	7	9	7	1	1	2	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.....

DATE 27/12/2022

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.