

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME CALPROTECTIN - CHORUS

Code	86600				
Lot No.	895		2023-12		2022-12

REAGENTS

Label	Description	Lot.	Exp.	NO
D9500	CALPROTECTIN	895	2023-12	2 X 6
B9500	CALIBRATORE CALPROTECTIN	895	2024-05	1
C9500	CONTROL CALPROTECTIN	895	2024-05	1
G9502	SAMPLE DILUENT	895	2024-12	1

CONTROLS


	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	350	≥ 100
Control (mg/kg)	126,3	78,6-177,3

CALIBRATOR

9	0	0	8	9	5	6	2	1	7	2	2	5	2	3	3	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader YES ☒ Operator Signature



CONTROL SERUM

9	0	0	8	9	5	6	1	4	2	5	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature



This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.....

DATE 29/12/2022

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent reagents should be handled with maximum attention.