

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME BETA 2-GLYCOPROTEIN-G - CHORUS

Code 86050

Lot No. 874



2023-09



2022-09

REAGENTS

Label	Description	Lot.	Exp.	NO
D3351	BETA 2-GLYCOPROTEIN-G	874	2023-09	6X6
B3351	CALIBRATORE BETA 2-GLYCOPROTEIN-G	874	2023-09	1
C3351	CONTROL SERUM BETA 2-GLYCOPROTEIN-G	874	2023-09	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	351	≥ 100
Control Serum (AU/mL)	29,1	22,6-40,9

CALIBRATOR


2	0	6	8	7	4	6	1	3	3	4	7	4	8	6	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature *88*



CONTROL SERUM


2	0	6	8	7	4	6	1	5	2	7	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature *88*



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... *Flavia*

DATE 29/12/2022

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.