

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI) – Italia – Via delle Rose 10
Tel +39-0577-587111 Fax +39-0577-318690



Sede Legale e Amministrazione
20144 Milano (Italy) Via A. Solari 19 sc. 6
Tel. +39-02-4859121 r.a. Fax +39-02-48008530

QUALITY CONTROL CERTIFICATE

KIT NAME: PSA - CHORUS

Code	86550				
Lot No.	887		2024-02		2022-11

REAGENTS

Label	Description	Lot.	Exp	No
D2020	PSA	887	2024-02	6X6
B2020	CALIBRATORE PSA	887	2024-10	1
C2020	CONTROL SERUM PSA	887	2024-10	1

CONTROLS

	Found results	Expected results
CALIBRATOR (mOD)	332	≥100
CONTROL SERUM (ng/ml)	16,8	14,5-20,6

CALIBRATOR


9	4	8	8	8	7	7	1	4	8	5	6	2	9	6	7	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature



CONTROL SERUM


9	4	8	8	8	7	7	2	9	4	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature



The test has been carried out in accordance with the package insert

TESTED BY: Longo Elvira

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.

DATE 02.02.2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested by an FDA approved method and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.