

DIESSE DIAGNOSTICA SENESE SpA
 53035 Monteriggioni (SIENA) – Italy – Strada dei Laghi, 39
 Tel. +39 0577319554 FAX +39 0577319020



Sede Legale e Amministrazione
 20144 Milano (Italy) Via A. Solari 19 sc. 6

QUALITY CONTROL CERTIFICATE

KIT NAME ENZYWELL CYTOMEGALOVIRUS IgG AVIDITY

Code 91092

LOT.NO 068 EXPIRY 2024-03

REAGENTS

CODE	LABEL	DESCRIPTION	LOT	EXP	NO
PF93010	007500	MICROPLATE STRIPS	260	2024-04	3
PF93501	007490	CONJUGATE	222	2024-03	1
PF91097	007470	LOW AVIDITY CONTROL	064	2024-06	1
PF91096	007480	HIGH AVIDITY CONTROL	124	2024-06	1
PF91091	007590	AVIDITY BUFFER	127	2024-06	1
PF93619	1034	SUBSTRATE	368	2024-04	1
PF93611	009660	DILUENT 2	226	2024-04	1
PF93603	1035	WASHING BUFFER 10x	453	2024-04	1
PF93602	1032/RISK0	STOP SOLUTION	215-1	2024-08	1

CONTROLS

	FOUND RESULTS	EXPECTED RESULTS
LOW AVIDITY CONTROL	11%	< 35%
HIGH AVIDITY CONTROL	67%	> 50%

The test has been carried out in accordance with the package insert

TESTED BY: Katia Tanzini

APPROVED BY: (RESPONSIBLE HEAD, Quality Control Reagents)

DATE: 11.01.23

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested by an FDA approved method and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.

**CERTIFICATE OF ANALYSIS**


KIT: CYTOMEGALOVIRUS IgG AVIDITY Code 91092
LOT: 068
DATE OF TEST: 09/01/2023

METHOD: ACCORDING TO THE PACKAGE INSERT

FOUND ***EXPECTED***

1. CONTROL OF COMPLETE KIT

- | | | | |
|----|---|-----|-------|
| a) | Low Avidity Control | 11% | < 35% |
| | High Avidity Control | 67% | > 50% |
| b) | Good function
Panel of Human Sera and Internal
Standards in accordance with
specifications | | |



SATISFACTORY



UNSATISFACTORY

FINAL RESULT: **ACCEPTED**

Date: 09.01.2023

Signed: 
Released: Q.C.Manager

Aggiornamento: 04/99