

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME CHORUS VIT B12

Code 86400

Lot No. 896



2023-12



2022-12

REAGENTS

Label	Description	Lot.	Exp	No
D4570	VIT B12	896	2023-12	6X6
B4570	Calibratore VIT B12	896	2025-06	1
C4570	Control VIT B12	896	2025-06	1
G4570	VIT B12 Sample Diluent/Negative Control	896	2025-06	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	340	≥ 100
Control (pg/ml)	432,9	275,0-520,0
VIT B12 Sample Diluent/Negative Control (pg/ml)	<150	< 150

CALIBRATOR


9	9	0	8	9	6	6	6	1	6	2	6	3	5	4	2	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature *AB*



CONTROL SERUM


9	9	0	8	9	6	6	2	3	3	4	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature *AB*



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... *Alvino*

DATE *20/01/2023*

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.