



DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020

Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

QUALITY CONTROL CERTIFICATE

KIT NAME CALPROTECTIN K - CHORUS

Code 86601

Lot No. 896



2023-12



2022-12

REAGENTS

Label	Description	Lot.	Exp.	NO
D9970	CALPROTECTIN	896	2023-12	2X6
B9970	CALIBRATORE CALPROTECTIN	896	2024-05	1
C9970	CONTROL CALPROTECTIN	896	2024-05	1
G9502	SAMPLE DILUENT	896	2024-12	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	270	≥ 100
Control Serum (mg/Kg)	414,3	244,4-623,8


CALIBRATOR

9	2	9	8	9	6	6	3	2	8	4	9	2	9	5	6	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES ☒

Operator Signature *F*




CONTROL SERUM

9	2	9	8	9	6	6	2	4	4	5	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES ☒

Operator Signature *F*



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *[Signature]*

DATE 13/6/2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.