

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME CHORUS PROMONITOR VEDOLIZUMAB

Code 86710

Lot No. 860-A



2023-08



2022-08

REAGENTS

Label	Description	Lot.	Exp.	NO
D4480	PROMONITOR VEDOLIZUMAB	860	2023-08	6x6
B4480	CALIBRATORE VEDOLIZUMAB	860	2023-12	1
C4480	CONTROL SERUM VEDOLIZUMAB	860	2023-12	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	463	≥ 100
Control Serum (ug/ml)	6,2	5,5-7,4

CALIBRATOR

2	7	8	8	6	0	6	8	1	4	3	0	3	1	8	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader

YES



Operator Signature

[Signature]



CONTROL SERUM

2	7	8	8	6	0	6	2	6	3	9	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader

YES



Operator Signature

[Signature]



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.....

[Signature]

DATE

18.01.2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.