



DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020

Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

QUALITY CONTROL CERTIFICATE

KIT NAME tTg-G - CHORUS

Code	86060		
Lot No.	884	2023-10	2022-10

REAGENTS

Label	Description	Lot.	Exp.	NO
D3451	tTg-G	884	2023-10	6X6
B3451	CALIBRATORE tTg-G	884	2023-10	1
C3451	CONTROL SERUM tTg-G	884	2023-10	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	465	≥ 100
Control Serum (AU/mL)	37,6	26,2-64,2

CALIBRATOR

3	4	0	8	8	4	6	1	1	4	2	2	5	5	4	1	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature

340884611422654100
tTg-G

CALIBRATOR 0.175 mL IVD

LOT 884

2023-10 2022-10

Monteriggioni Italy

CONTROL SERUM

3	4	0	8	8	4	6	1	3	2	4	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature

340884613240000001
tTg-G

CONTROL 0.425 mL IVD

LOT 884

2023-10 2022-10

Monteriggioni Italy

*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.....

DATE 16/01/2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent reagents should be handled with maximum attention.