

DIESE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI) – Italia – Via delle Rose 10
Tel +39-0577-587111 Fax +39-0577-318690



Pag 1/1
Sede Legale e Amministrazione
20144 Milano (Italy) Via A. Solari 19 sc. 6
Tel. +39-02-4859121 r.a. Fax +39-02-48008530

QUALITY CONTROL CERTIFICATE

KIT NAME: 88000- G15 - CULTIVATED WHEAT

Code	88000 - G15			
Lot No.	890	2023-11	2023-01	

REAGENTS

Label	Description	Lot.	Exp	No
D5051	88000- G15 CULTIVATED WHEAT	890	2023-11	2X6
B5050	CALIBRATORE 88000- G15 CULTIVATED WHEAT	890	2024-02	1
C5050	CONTROL SERUM 88000- G15 CULTIVATED WHEAT	890	2024-02	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	232	≥ 100
Control Serum (KU/L)	13,0	8,8-20,3

CALIBRATOR


6	8	4	8	9	0	6	5	1	7	5	3	2	3	9	5	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature *SB*



CONTROL SERUM


6	8	4	8	9	0	6	3	0	5	6	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature *SB*



The test has been carried out in accordance with the package insert

TESTED BY: Simone Baiocchi

Simone Baiocchi

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

[Signature]

DATE

16.2.2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested by an FDA approved method and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.