

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI) – Italia – Via delle Rose 10
Tel +39-0577-587111 Fax +39-0577-318690



Pag 1/1
Sede Legale e Amministrazione
20144 Milano (Italy) Via A. Solari 19 sc. 6
Tel. +39-02-4859121 r.a. Fax +39-02-48008530

QUALITY CONTROL CERTIFICATE

KIT NAME: 88000-F76 ALPHA-LACTALBUMIN

Code	88000-F76		
Lot No.	890	2023-11	2023-01

REAGENTS


Label	Description	Lot	Exp	No
D9861	88000-F76 ALPHA-LACTALBUMIN	890	2023-11	2X6
B9860	CALIBRATORE 88000-F76 ALPHA-LACTALBUMIN	890	2024-02	1
C9860	CONTROL SERUM 88000-F76 ALPHA-LACTALBUMIN	890	2024-02	1

CONTROLS

	Found results	Expected result
Calibrator (mOD)	232	≥ 100
Control Serum (KU/L)	13,0	8,8-20,3


CALIBRATOR

6	1	9	8	9	0	6	5	1	7	5	3	2	3	9	5	0	0	
Label checked with barcode reader										YES	☒	Operator Signature <i>SB</i>						



CONTROL SERUM

6	1	9	8	9	0	6	3	0	5	6	0	0	0	0	0	0	1	
Label checked with barcode reader										YES	☒	Operator Signature <i>SB</i>						



The test has been carried out in accordance with the package insert

TESTED BY: Simone Baiocchi *Simone Baiocchi*
APPROVED RESPONSIBLE HEAD, Quality Control Reagents *[Signature]*
DATE 19.01.23

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested by an FDA approved method and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.