

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME CARDIOLIPIN-M - CHORUS

Code 86048

Lot No. 897



2023-12



2022-12

REAGENTS

Label	Description	Lot.	Exp.	NO
D3331	CARDIOLIPIN-M	897	2023-12	6X6
B3331	CALIBRATORE CARDIOLIPIN-M	897	2023-12	1
C3331	CONTROL SERUM CARDIOLIPIN-M	897	2023-12	1

CONTROLS


	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	377	≥ 100
Control Serum (MPLU/mL)	51,8	36,7-90,8

CALIBRATOR

3	3	6	8	9	7	6	4	1	7	2	6	3	7	3	9	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader YES ☒ Operator Signature *SB*



CONTROL SERUM

3	3	6	8	9	7	6	1	9	3	6	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature *SB*



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents. *SB*

DATE 23.01.2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.