

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SIENA) – Italy – Strada dei Laghi, 39
Tel. +39 0577319554 FAX +39 0577319020

DIESSE

Sede Legale e Amministrazione
20144 Milano (Italy) Via A. Solari 19 sc. 6

QUALITY CONTROL CERTIFICATE

KIT NAME: AMA-M2 - CHORUS

Code	86084			
Lot No.	887		2024-02	 2022-12

REAGENTS

Label	Description	Lot.	Exp	No
D3651	AMA-M2	887	2024-02	6X6
B3651	CALIBRATORE AMA-M2	887	2024-02	1
C3651	CONTROL SERUM AMA-M2	887	2024-02	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	398	≥ 100
Control Serum (AU/mL)	60,0	32,5-100,0

CALIBRATOR


7	0	5	8	8	7	7	4	2	4	3	2	4	2	4	8	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature 



CONTROL SERUM


7	0	5	8	8	7	7	1	7	3	2	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature 



The test has been carried out in accordance with the package insert

TESTED BY: Longo Elvira 

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... 

DATE 27.01.2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested by an FDA approved method and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.