

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI) – Italia – Via delle Rose 10
Tel +39-0577-587111 Fax +39-0577-318690



Pag 1/1
Sede Legale e Amministrazione
20144 Milano (Italy) Via A. Solari 19 sc. 6
Tel. +39-02-4859121 r.a. Fax +39-02-48008530

QUALITY CONTROL CERTIFICATE

KIT NAME: FSH

Code	86510		
Lot No.	895	2023-09	2023-01

REAGENTS

Label	Description	Lot.	Exp	No
D3230	FSH	895	2023-09	6X6
B3230	CALIBRATOR FSH	895	2023-09	1
C3230	CONTROL SERUM FSH	895	2023-09	1

CONTROLS

	Found results	Expected results
CALIBRATOR (mOD)	269	≥100
CONTROL SERUM (mIU/ml)	12,4	8,5-16,7

CALIBRATOR


9	2	0	8	9	5	5	5	1	8	4	7	2	8	5	3	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature *SB*



CONTROL SERUM


9	2	0	8	9	5	5	2	6	4	8	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature *SB*



The test has been carried out in accordance with the package insert.

TESTED BY: Simone Balocchi

APPROVED

RESPONSIBLE HEAD, Quality Control Reagents

DATE

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.