

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020

DIESSE

Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME CALPROTECTIN - CHORUS

Code 86600/C03

Lot No. 897 2023-12 2022-12

REAGENTS

Label	Description	Lot.	Exp.	NO
D9500	CALPROTECTIN	897	2023-12	2X6
D9510	CALPROTECTIN	897	2023-12	1X4
B9510	CALIBRATORE CALPROTECTIN	897	2024-05	1
C9510	CONTROL CALPROTECTIN	897	2024-05	1
G9502	SAMPLE DILUENT	897	2024-12	1

CONTROLS


	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	442	≥ 100
Control (mg/kg)	104,7	65,0-161,1

CALIBRATOR

9 0 0 8 9 7 6 2 2 0 2 4 5 2 3 3 0 0

Label checked with barcode reader YES ☒ Operator Signature *[Signature]*



CONTROL SERUM

9 0 0 8 9 7 6 1 3 2 5 0 0 0 0 0 0 1

Label checked with barcode reader YES ☒ Operator Signature *[Signature]*



This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... *[Signature]*

DATE 24/01/2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.