

DIESSE DIAGNOSTICA SENESE SpA  
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39  
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale  
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

## QUALITY CONTROL CERTIFICATE

KIT NAME BETA 2-GLYCOPROTEIN-M - CHORUS

|         |       |         |         |
|---------|-------|---------|---------|
| Code    | 86052 |         |         |
| Lot No. | 895   | 2023-12 | 2022-12 |

### REAGENTS

| Label | Description                         | Lot. | Exp.    | NO  |
|-------|-------------------------------------|------|---------|-----|
| D3371 | BETA 2-GLYCOPROTEIN-M               | 895  | 2023-12 | 6X6 |
| B3371 | CALIBRATORE BETA 2-GLYCOPROTEIN-M   | 895  | 2023-12 | 1   |
| C3371 | CONTROL SERUM BETA 2-GLYCOPROTEIN-M | 895  | 2023-12 | 1   |


### CONTROLS

|                       | Found results | Expected results |
|-----------------------|---------------|------------------|
| Calibrator (mOD)      | 298           | > 100            |
| Control Serum (AU/mL) | 60,7          | 42,6-90,9        |

### CALIBRATOR

|   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|
| 2 | 0 | 5 | 8 | 9 | 5 | 6 | 4 | 2 | 9 | 5 | 1 | 2 | 8 | 7 | 7 | 0 | 0 |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|

Label checked with barcode reader YES ☒ Operator Signature .....



### CONTROL SERUM

|   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|
| 2 | 0 | 5 | 8 | 9 | 5 | 6 | 2 | 5 | 4 | 7 | 0 | 0 | 0 | 0 | 0 | 0 | 1 |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|

Label checked with barcode reader YES ☒ Operator Signature .....



*This batch has been manufactured and checked according to quality requirements.  
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.....

DATE 25.01.2023

### WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.