### DIESSE DIAGNOSTICA SENESE SpA

53035 Monteriggioni (SI), Italy, Strada dei Laghi 39 Tel. +39 0577319554 FAX +39 0577319020



Sede Legale 20144 Milano (Italy) Via A. Solari 19 sc. 6

# Code 86713 Lot No. 895 QUALITY CONTROL CERTIFICATE M. 2022-12

### REAGENTS

bel	Description	Lot.	Exp.	NO
D4510	PROMONITOR ANTI-USTEKINUMAB	895	2023-12	6x6
B4510	CALIBRATORE ANTI-USTEKINUMAB	895	2024-10	1
C4510	CONTROL SERUM ANTI-USTEKINUMAB	895	2024-10	1

### CONTROLS

Found results	Expected results
I our a results	Expected results

Calibrator (mOD)	300	≥ 100
Control Serum (AU/mL)	23,1	15,4-26,5

## CALIBRATOR

2 8 1 8 9 5 6 8 2 1 3 4 2 9 5 5 0 0


Label checked with barcode reader

YES 🔀

Operator Signature



# CONTROL SERUM



Label checked with barcode reader

YES X

Operator Signature



This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.

DATE 26-01 623

### WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg, Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.

IO 09-456-C cod 86713