

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME 88000-F13 PEANUT

Code 88000-F13

Lot No. 904



2024-01



2023-01

REAGENTS

| Label | Description | Lot. | Exp. | NO |
|-------|--------------------------------|------|---------|-----|
| D9941 | 88000-F13 PEANUT | 904 | 2024-01 | 2x6 |
| B9940 | CALIBRATORE 88000-F13 PEANUT | 904 | 2024-02 | 1 |
| C9940 | CONTROL SERUM 88000-F13 PEANUT | 904 | 2024-02 | 1 |

CONTROLS

| | Found results | Expected results |
|--|---------------|------------------|
|--|---------------|------------------|

| | | |
|----------------------|------|------------|
| Calibrator (mOD) | 260 | ≥ 100 |
| Control Serum (KU/L) | 14,0 | 7,9-17,6 |

CALIBRATOR

| | | | | | | | | | | | | | | | | | |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|
| 6 | 1 | 7 | 9 | 0 | 4 | 6 | 5 | 1 | 7 | 5 | 3 | 2 | 3 | 9 | 5 | 0 | 0 |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|

Label checked with barcode reader

YES



Operator Signature

SB



CONTROL SERUM

| | | | | | | | | | | | | | | | | | |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|
| 6 | 1 | 7 | 9 | 0 | 4 | 6 | 3 | 0 | 5 | 6 | 0 | 0 | 0 | 0 | 0 | 0 | 1 |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|


Label checked with barcode reader

YES



Operator Signature

SB



This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

SB

DATE

31/01/2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.