

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME tTg-A - CHORUS

Code 86058

Lot No. 898



2023-12



2022-12

REAGENTS

Label	Description	Lot.	Exp.	NO
D3431	tTg-A	898	2023-12	6X6
B3431	CALIBRATORE tTg-A	898	2023-12	1
C3431	CONTROL SERUM tTg-A	898	2023-12	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	408	≥ 100
Control Serum (AU/mL)	41,2	27,0-60,0

CALIBRATOR

3	3	9	8	9	8	6	2	2	0	3	4	3	8	4	3	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader

YES



Operator Signature

RB



CONTROL SERUM

3	3	9	8	9	8	6	1	8	3	4	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader

YES



Operator Signature

RB



This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

RB

DATE

30.01.2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent reagents should be handled with maximum attention.