

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020

DIESSE

Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME HSV 2 SCREEN - CHORUS

Code	81019				
Lot No.	902		2024-04		2023-01

REAGENTS

Label	Description	Lot.	Exp.	NO
D2201	HSV 2 SCREEN	902	2024-04	6x6
B2201	CALIBRATORE HSV 2 SCREEN	902	2024-05	1
C2201	CONTROL SERUM HSV 2 SCREEN	902	2024-08	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	520	≥ 100
Control Serum (IU/mL)	1,4	1,3-8,0
Negative Control (IU/mL)	0,6	$< 0,7$

CALIBRATOR

0	3	4	9	0	2	7	0	0	0	0	0	6	8	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature 



CONTROL SERUM

0	3	4	9	0	2	7	1	3	8	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature 



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents 

DATE 06.02.2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.