

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME tTg-G - CHORUS

Code	86060		
Lot No.	897	2023-12	2022-12

REAGENTS

Label	Description	Lot.	Exp.	NO
D3451	tTg-G	897	2023-12	6X6
B3451	CALIBRATORE tTg-G	897	2023-12	1
C3451	CONTROL SERUM tTg-G	897	2023-12	1

CONTROLS


	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	448	≥ 100
Control Serum (AU/mL)	39,7	34,0-64,4

CALIBRATOR

3	4	0	8	9	7	6	1	1	5	3	0	4	2	3	9	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader YES ☒ Operator Signature



CONTROL SERUM

3	4	0	8	9	7	6	1	7	3	1	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

DATE 02.01.2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent reagents should be handled with maximum attention.