

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

QUALITY CONTROL CERTIFICATE

KIT NAME Scl-70 - CHORUS

| | | | |
|---------|-------|---------|---------|
| Code | 86026 | | |
| Lot No. | 896 | 2023-12 | 2022-12 |

REAGENTS

| Label | Description | Lot. | Exp. | NO |
|-------|----------------------|------|---------|-----|
| D3221 | Scl-70 | 896 | 2023-12 | 6X6 |
| B3221 | CALIBRATORE Scl-70 | 896 | 2023-12 | 1 |
| C3221 | CONTROL SERUM Scl-70 | 896 | 2023-12 | 1 |

CONTROLS

| | Found results | Expected results |
|--|---------------|------------------|
|--|---------------|------------------|

| | | |
|-----------------------|------|-----------|
| Calibrator (mOD) | 445 | ≥ 100 |
| Control Serum (AU/mL) | 34,1 | 25,9-66,2 |

CALIBRATOR

| | | | | | | | | | | | | | | | | | |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|
| 1 | 9 | 5 | 8 | 9 | 6 | 6 | 2 | 2 | 1 | 3 | 2 | 3 | 8 | 4 | 5 | 0 | 0 |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|


Label checked with barcode reader YES ☒ Operator Signature



CONTROL SERUM

| | | | | | | | | | | | | | | | | | |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|
| 1 | 9 | 5 | 8 | 9 | 6 | 6 | 1 | 8 | 3 | 4 | 0 | 0 | 0 | 0 | 0 | 0 | 1 |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|

Label checked with barcode reader YES ☒ Operator Signature



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

DATE 06/02/2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.