

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI) – Italia – Strada dei Laghi 39
Tel +39-0577-587111 Fax +39-0577-318690



Pag 1/1
Sede Legale e Amministrazione
20144 Milano (Italy) Via A. Solari 19 sc. 6
Tel. +39-02-4859121 r.a. Fax +39-02-48008530

QUALITY CONTROL CERTIFICATE

KIT NAME: FT4 - CHORUS

Code	86520		
Lot No.	900	2024-01	2023-01

REAGENTS

Label	Description	Lot.	Exp	No
D3320	FT4	900	2024-01	6X6
B3320	CALIBRATOR FT4	900	2024-01	1
C3320	CONTROL SERUM FT4	900	2024-01	1

CONTROLS

Found results	Expected results
---------------	------------------


CALIBRATOR (mOD)	351	≥100
CONTROL SERUM (pmol/L)	21,8	2,6-32,5

CALIBRATOR

9	6	8	9	0	0	6	2	2	7	3	4	3	3	3	9	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES

Operator Signature *SB*

CONTROL SERUM

9	6	8	9	0	0	6	2	1	3	9	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES

Operator Signature *SB*

The test has been carried out in accordance with the package insert

TESTED BY: Simone Baiocchi *Simone Baiocchi*APPROVED RESPONSIBLE HEAD, Quality Control Reagents *Baiocchi*

DATE 14.02.2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.