

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI) – Italia – Strada dei Laghi 39
Tel +39-0577-587111 Fax +39-0577-318690



Sede Legale e Amministrazione
20144 Milano (Italy) Via A. Solari 19 sc. 6
Tel. +39-02-4859121 r.a. Fax +39-02-48008530

QUALITY CONTROL CERTIFICATE

KIT NAME: 88000-C436 - SPIRAMYCIN

Code 88000 - C436			
Lot No. 904	2024-01	2023-02	

REAGENTS

Label	Description	Lot.	Exp	No
D1800	88000- C436 SPIRAMYCIN	904	2024-01	2X6
B1800	CALIBRATORE 88000- C436 SPIRAMYCIN	904	2024-02	1
C1800	CONTROL SERUM 88000- C436 SPIRAMYCIN	904	2024-02	1

CONTROLS


	Found results	Expected results
Calibrator (mOD)	260	≥ 100
Control Serum (KU/L)	14,0	7,9-17,6

CALIBRATOR

7	8	3	9	0	4	6	5	1	7	5	3	2	3	9	5	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES


Operator Signature *RB*

CONTROL SERUM

7	8	3	9	0	4	6	3	0	5	6	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES

Operator Signature *RB*

The test has been carried out in accordance with the package insert

TESTED BY: Simone Balocchi

APPROVED

RESPONSIBLE HEAD, Quality Control Reagents

DATE

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested by an FDA approved method and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.