

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI) – Italia – Via delle Rose 10
Tel +39-0577-587111 Fax +39-0577-318690



Sede Legale e Amministrazione
20144 Milano (Italy) Via A. Solari 19 sc. 6
Tel. +39-02-4859121 r.a. Fax +39-02-48008530

QUALITY CONTROL CERTIFICATE

KIT NAME: TSH

Code	86516			
Lot No.	901		2024-01	2023-01

REAGENTS

Label	Description	Lot.	Exp	No
D3290	TSH	901	2024-01	6X6
B3290	CALIBRATOR TSH	901	2024-01	1
C3290	CONTROL SERUM TSH	901	2024-01	1

CONTROLS

Found results	Expected results
---------------	------------------

CALIBRATOR (mOD)	405	≥100
CONTROL SERUM (uIU/ml)	3,6	2,1-6,1

CALIBRATOR


9	1	9	9	0	1	6	7	1	4	1	8	4	6	4	6	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature *f*



CONTROL SERUM


9	1	9	9	0	1	6	1	5	2	9	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature *f*



The test has been carried out in accordance with the package insert

TESTED BY: LONGO ELVIRA *f*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... *f*

DATE 09/02/2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.