

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME dsDNA-G - CHORUS

Code 86032

Lot No. 895



2023-12



2022-12

REAGENTS

Label	Description	Lot.	Exp.	NO
D3281	dsDNA-G	895	2023-12	6X6
B3281	CALIBRATORE dsDNA-G	895	2023-12	1
C3281	CONTROL SERUM dsDNA-G	895	2023-12	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	400	≥ 100
Control Serum (IU/mL)	54,6	35,0-85,0

CALIBRATOR

2	5	1	8	9	5	6	7	1	5	2	5	6	1	2	8	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader YES ☒ Operator Signature



CONTROL SERUM

2	5	1	8	9	5	6	1	2	2	2	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.....

DATE 09.02.2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.